

Name: Wuxi Institute of Inspection, Testing and Certification(Wuxi Institute of Metrology and Testing/Wuxi Center of Fiber Inspection)

Address: No.200, Linghu Avenue, Xinwu District, Wuxi, Jiangsu, China

Registration No. CNAS L0260

Accreditation Criteria: ISO/IEC 17025:2017 and relevant requirements of CNAS

Effective Date: 2025-06-10 Expiry Date: 2028-09-16

SCHEDULE 5 ACCREDITED CALIBRATION AND MEASUREMENT CAPABILITY SCOPE

Note: The instruments with * represents onsite calibration can be performed.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
I、Mechanical measuring instrument							
1	Vibration Displacement Transducer	Displacement	Verification Regulation of Vibration Displacement Transducer JJG 644	Static:(0.01~25)mm	$U_{rel}=0.3\%$		
				Dynamic:(0.01~5)mm, (0.4~500)Hz	$U_{rel}=1.5\%$		
2	Linear Accelerometer	Linear acceleration	Calibration Specification for MEMS Linear Accelerometers JJF1427	(9.8~450)m/s ² , (32~330)r/min	$U_{rel}=5 \times 10^{-4}$		
				(-9.8~-1.7×10 ⁻⁴)m/s ² , (-55~85)°C	$U_{rel}=2 \times 10^{-5}$		
				(1.7×10 ⁻⁴ ~9.8)m/s ² , (-55~85)°C	$U_{rel}=2 \times 10^{-5}$		
3	Vibratory Transmitter	Acceleration	Calibration Specification for Vibratory Transmitter JJF(Su)235	(0.1~100)m/s ² , (0.4~4000)Hz	$U_{rel}=1.5\%$		



No. CNAS L0260

第 1 页 共 8 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Velocity		(0.1~100)mm/s, (0.4~1000)Hz	$U_{rel}=1.5\%$		
		Displacement		(0.01~5)mm, (0.4~500)Hz	$U_{rel}=1.5\%$		
4	Electromagnetic Sensor of Rotational Speed	Rotational Speed	C. S. for Electromagnetic Sensor of Rotational Speed JJF1871	(10~100)r/min	$U=0.03r/min$		
				(100~8000)r/min	$U_{rel}=0.01\%$		
5	Speed and Milege Meter for Standard Equipment	Rotate Speed	V.R.of Speed and Milege Meter for Standard Equipment JJG779	(10~4000)r/min	$U_{rel}=0.6 \times 10^{-4}$		
6	Tachometer	Rotate Speed	V.R.of Tachometers JJG105	(10~40000)r/min	$U_{rel}=0.02\%$		
7	Aneroid Barograph and Aneroid Barometer	Pressure	V.R.of Aneroid Barograph and Aneroid Barometer JJG272	(500~1050)hPa	$U=0.4hPa$		
8	Equipment Standard for Revolution Speed	Rotate Speed	V.R.of Standard Equipment for Revolution Speed JJG326	(10~40000)r/min	$U_{rel}=0.9 \times 10^{-4}(k=3)$		
9	Digital Barometers	Pressure	V.R.of Digital Barometers JJG1084	(10~1200)hPa	$U=(0.2 \sim 0.4)hPa$		
10	Accelerometer	Acceleration	V.R.of Piezoelectric Accelerometer JJG233	(0.1~100)m/s ² , (0.4~20)Hz	$U_{rel}=1.1\%$		
				(0.1~1000)m/s ² , (20~10000)Hz	$U_{rel}=1.1\%$		
11	Measuring Vibration Instruments	Acceleration	V.R.of Vibration Meters JJG676	(0.1~100)m/s ² , (5~2000)Hz	$U_{rel}=2.3\%$		
		Velocity		(0.1~100)mm/s, (5~1000)Hz	$U_{rel}=2.3\%$		
		Displacement		(0.01~5)mm, (5~500)Hz	$U_{rel}=2.3\%$		



No. CNAS L0260

第 2 页 共 8 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Frequency		(5~2000)Hz	$U_{rel}=0.2\%$		
12	Portable Vibration Calibrator	Acceleration	V.R.of Portable Vibration Calibrator JJG 1062	(0.1~1000)m/s ² , (5~ 4000) Hz	$U_{rel}=1.2\%$		
		Frequency		(5~4000)Hz	$U_{rel}=0.06\%$		
		Distortion		0.01%~30%	$U_{rel}=2\%$		
13	Photoelectric Belt Tension Meters	Frequency	Calibration Specification for Photoelectric Belt Tension Meters JJF(S) 204	(10~50)Hz	$U=0.1\text{Hz}$		
				(50~600) Hz	$U=0.2\text{Hz}$		
14	Electromagnetic Velocity Transducers	Velocity	Electromagnetic Velocity Transducers JJG 134	(0.1~100)mm/s, (5~ 4000)Hz	$U_{rel}=1.1\%$		
15	Capacitance Accelerometers	Acceleration	Calibration Specification for Capacitance Accelerometers JJF 1918	Static: (0~9.8) m/s ² , (0~180) °	$U=0.005\text{m/s}^2$		
				Dynamic:(0.1~ 100)m/s ² , (0.4~20)Hz	$U_{rel}=1.1\%$		
				Dynamic:(0.1~ 1000)m/s ² , (20~ 10000)Hz	$U_{rel}=1.1\%$		
16	Linear Velocity Measuring Instrument	Linear Velocity	Calibration Specification for Linear Velocity Measuring Instrument JJF 1801	contact: (0.1~4) m/s	$U_{rel}=0.6\%$		
				two points: (0.3~30) m/s	$U_{rel}=0.3\%$		
17	Human Vibration Meters	Weighted acceleration	Verification Regulation of Human Vibration Meters JJG 1178	Frequency: (0.4~ 1000)Hz; Acceleration: (0.01~100) m/s ²	$U_{rel}=1.1\%$		
18		Acceleration	Calibration Specification for Shock Measuring Instruments JJF 1943	Acceleration: (200~ 50000)m/s ² ; Pulse Duration:(0.05~10)ms	$U_{rel}=4.3\%$		



No. CNAS L0260

第 3 页 共 8 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
19		overload Acceleration	Calibration Specification for Acceleration Overload Sensors JJF 2038	Acceleration: (8~ 490)m/s ² , (20~2000) Hz	$U_{rel}=0.4\%$		
II、Acoustic measuring instrument							
1	Personal Sound Exposure Meter	sound exposure	Personal Sound Exposure Meters JJG 980	Absolute acoustic sensitivity:10Pa ² h	$U=0.9\text{Pa}^2\text{h}$		
				Weighted, Steady State Response:(0.1~ 99.99)Pa ² h, (63~ 8000)Hz	$U_{rel}=6\%$		
				Short duration signal response:(0.1~ 99.99)Pa ² h, (63~ 8000)Hz	$U=0.12\text{Pa}^2\text{h}$		
2	Microphone Preamplifier	Frequency Response	Calibration Specification for Microphone Preamplifiers JJF1137	(-10~10)dB, (10 Hz~ 50 kHz)	$U=0.3\text{dB}$		
		Transmission loss		(-10~10)dB, (10 Hz~ 50 kHz)	$U=0.2\text{dB}$		
3	Dynamical Signal Analyzer	Frequency	V.R. of Dynamical Signal Analyzer JJG834	0.1Hz~200kHz	$U_{rel}=6\times 10^{-5}$		
		Voltage		1mV~10V	$U_{rel}=0.6\%$		
4	Pure tone audiometer	Hearing level zero	V.R. of Audiological Equipment Pure-tone Audiometers JJG388	Air conduction : (-10~ 120) dB, 125Hz~8kHz	$U=0.9\text{dB}$		
				Bone conduction : (- 10~70) dB, 125Hz~ 8kHz	$U=1.7\text{dB}$		
5	Sound level meter	Sound Pressure Level	V.R. of Sound Level Meters JJG188	Sound Signal:(30~ 130)dB, (10Hz~4kHz)	$U=0.6\text{dB}$		



No. CNAS L0260

第 4 页 共 8 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				Sound Signal:(30~130)dB, (4~10)kHz)	<i>U</i> =0.7dB		
				Sound Signal:(30~130)dB, (10~20)kHz	<i>U</i> =1.0dB		
				Sine Signal:(30~130)dB, (31.5Hz~12.5kHz)	<i>U</i> =0.2dB		
				Toneburst Signal:(30~130)dB, (0.25~1000)ms	<i>U</i> =0.3dB		
		Time-Weighting F and S	F:(20.0~50.0)dB/s	<i>U</i> =3.2dB/s			
			S:(1.0~8.0)dB/s	<i>U</i> =0.3dB/s			
		6	Noise Level Statistical Analyzers	Sound Pressure Level	V.R.of Noise Level Statistical Analyzers JJG778		Sound Signal:(30~130)dB, (10Hz~4kHz)
Sound Signal:(30~130)dB,(4kHz~10kHz)	<i>U</i> =0.7dB						
Sound Signal: (0~130)dB, (10~20)kHz	<i>U</i> =1.0dB						
Sine Signal: (30~130)dB, (31.5Hz~12.5kHz)	<i>U</i> =0.2dB						
Toneburst Signal:(30~130)dB, (0.25~1000)ms	<i>U</i> =0.3dB						
Time-Weighting F and S	F:(20.0~50.0)dB/s			<i>U</i> =3.2dB/s			
	S:(1.0~8.0)dB/s			<i>U</i> =0.3dB/s			
7	Acoustic calibrator	SPL	V.R.of Acoustic calibrator JJG176	(80~130)dB,(20~8000)Hz	<i>U</i> =0.12 dB		



No. CNAS L0260

第 5 页 共 8 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Frequency		(20~8000)Hz	$U_{rel}=0.2\%$		
		Total Distortion		(0.03~10.0)%	$U=0.40\%$		
8	Working Standard Microphones(Electrostatic Actuator Method)	sound pressure sensitivity level	V.R.of Working Standard Microphones(Electrostatic Actuator Method) JJG 175	Frequency response of sound pressure sensitivity level: (10~140) dB, (20~20000) Hz	$U= (0.16\sim0.50)$ dB		
9	Working Standard Microphones(Coupler Comparison Method)	sound pressure sensitivity level	V.R.of Working Standard Microphones(Coupler Comparison Method) JJG 1019	Frequency response of sound pressure sensitivity level: (10~140) dB, (20~20000) Hz	$U= (0.14\sim0.50)$ dB		
10	Anechoic Rooms and Hemi-anechoic Rooms	sound pressure level	Calibration Specification for Acoustic Performance of Anechoic Rooms and Hemi-anechoic Rooms JJF 1147	Antisquare sound pressure level for Anechoic Rooms : 100Hz~20kHz	$U=0.7$ dB		
				Hemi-anechoic Rooms : 100Hz~20kHz	$U=1.0$ dB		
				Background Noise; (6.5~110) dB; 20Hz~20kHz	$U=0.7$ dB		
11	Multi-Channels Sound Analyzers	Sound Pressure Level	Calibration Specification for Multi-Channels Sound Analyzers JJF 1288	Frequency-Weighting: 20dB~140dB, (10Hz~20kHz)	$U=0.1$ dB		
				Tone Burst Response:20dB~140dB, (10Hz~20kHz)	$U=0.2$ dB		
				Repetitive Tone Burst Response:20dB~140dB, (10Hz~20kHz)	$U=0.2$ dB		



No. CNAS L0260

第 6 页 共 8 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date	
				Linear level: 20dB~140dB, (10Hz~20kHz)	U=0.2dB			
				Amplitude Frequency Characteristic: 20dB~140dB, (10Hz~20kHz)	U=0.1dB			
				Pink Noise and White Noise: 20dB~140dB, (20Hz~20kHz)	U=1.2dB			
		Decay Rate		F: 20.0dB/s~50.0dB/s	U=3.2dB/s			
				S: 2.0dB/s~8.0dB/s	U=0.3dB/s			
		Reverberation Time		0.75s、1.5s、7.5s、22.5s	U _{rel} =1.2%			
		Frequency		10Hz~20kHz	U _{rel} =0.04%			
	AC Voltage	10mV~10V, (10Hz~20kHz)	U _{rel} =0.1%					
III、Electromagnetic measuring instrument								
1	Fluxgate Magnetometer	Magnetic Induction	Calibration Specification for Fluxgate Magnetometer	(-100~100) μ T	U=0.3%R _d +10nT			
		Orthogonality	JJF1519	(0~10)°	U=0.15°			
Special measuring instruments								
IV、Special measuring instruments for construction and traffic								
1	Pile Dynamic Measuring Instrument	Acceleration	V.R.of Pile Dynamic	(0.5~1000)m/s ²	U _{rel} =1.4%			
		Strain	Measuring InstrumentJJG930-1998	(1~2000) μ ε	U _{rel} =2%			
		Frequency	JJG930	(2~5000)Hz	U _{rel} =0.2%			



No. CNAS L0260

第 7 页 共 8 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
V、Other measuring instruments							
1	MEMS Gyroscopes	Angular speed	Calibration Specification for MEMS Gyroscopes JJF 1535	(0.01~500)° /s, (-55~85)°C	<i>U</i> _{rel} =1.1×10 ⁻³		
		(500~1500)° /s		<i>U</i> _{rel} =1.1×10 ⁻³			
		Scale Factor		(1~1×10 ⁶)mV/(° /s)	<i>U</i> _{rel} =1.4×10 ⁻³		
VI、Geometric measuring instrument							
1		Angle	Calibration Specification for Single Axis Inclination Sensors JJF 2015	Angle:-45° ~+45°	<i>U</i> =0.02°		



No. CNAS L0260

第 8 页 共 8 页

The scope of the accreditation in Chinese remains the definitive version.